



DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

[Docket No. FDA-2016-N-0001]

Clinical Chemistry and Clinical Toxicology Devices Panel of the Medical Devices Advisory Committee; Notice of Meeting

AGENCY: Food and Drug Administration, HHS.

ACTION: Notice.

SUMMARY: The Food and Drug Administration (FDA) announces a forthcoming public advisory committee meeting of the Clinical Chemistry and Clinical Toxicology Devices Panel of the Medical Devices Advisory Committee. The general function of the committee is to provide advice and recommendations to the Agency on FDA's regulatory issues. The meeting will be open to the public.

DATES: The meeting will be held on August 10, 2016, from 8 a.m. to 6 p.m.

ADDRESSES: Gaithersburg Holiday Inn, Ballroom, Two Montgomery Village Ave., Gaithersburg, MD 20879. The hotel's telephone number is 301-948-8900. Answers to commonly asked questions including information regarding special accommodations due to a disability, visitor parking, and transportation may be accessed at:

<http://www.fda.gov/AdvisoryCommittees/AboutAdvisoryCommittees/ucm408555.htm>.

FOR FURTHER INFORMATION CONTACT: Patricio G. Garcia, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, rm. 1611, Silver Spring, MD 20993-0002, Patricio.Garcia@fda.hhs.gov, 301-796-6875, or FDA Advisory Committee Information Line, 1-800-741-8138 (301-443-0572 in the Washington, DC

area). A notice in the **Federal Register** about last minute modifications that impact a previously announced advisory committee meeting cannot always be published quickly enough to provide timely notice. Therefore, you should always check the Agency's Web site at <http://www.fda.gov/AdvisoryCommittees/default.htm> and scroll down to the appropriate advisory committee meeting link, or call the advisory committee information line to learn about possible modifications before coming to the meeting.

SUPPLEMENTARY INFORMATION:

Agenda: On August 10, 2016, the committee will discuss, make recommendations, and vote on information regarding a de novo request for the SEEKER Newborn Screening System (SEEKER System), by Baebies, Inc. The SEEKER System consists of the SEEKER Analyzer, the SEEKER 4-Plex Assay Kit, the SEEKER Cartridges, the Spot Logic software, and quality control materials; it uses digital microfluidic technology to measure multiple lysosomal enzymatic activities quantitatively from newborn dried blood spot specimens. The proposed Indication for Use for the SEEKER System device, as stated in the de novo request, is as follows:

The SEEKER System is intended for quantitative measurement of the activity of multiple lysosomal enzymes from newborn dried blood spot specimens. Reduced activity of these enzymes may be indicative of a lysosomal storage disorder. The enzymes measured using the SEEKER 4-Plex Assay Kit and their associated lysosomal storage disorder are listed in the following table.

| Enzyme (abbreviation) | Disorder |
|-----------------------------------|--|
| α -L-iduronidase (IDUA) | Mucopolysaccharidosis Type I (MPS I) disease |
| α -D-glucosidase (GAA) | Pompe disease |
| β -glucocerebrosidase (GBA) | Gaucher disease |
| α -D-galactosidase A (GLA) | Fabry disease |

Reduced activity for any of the four enzymes must be confirmed by other confirmatory diagnostic methods.

FDA intends to make background material available to the public no later than 2 business days before the meeting. If FDA is unable to post the background material on its Web site prior to the meeting, the background material will be made publicly available at the location of the advisory committee meeting, and the background material will be posted on FDA's Web site after the meeting. Background material is available at <http://www.fda.gov/AdvisoryCommittees/Calendar/default.htm>. Scroll down to the appropriate advisory committee meeting link.

Procedure: Interested persons may present data, information, or views, orally or in writing, on issues pending before the committee. Written submissions may be made to the contact person on or before August 3, 2016. On August 10, 2016, oral presentations from the public will be scheduled between approximately 1 p.m. and 2 p.m. Those individuals interested in making formal oral presentations should notify the contact person and submit a brief statement of the general nature of the evidence or arguments they wish to present, the names and addresses of proposed participants, and an indication of the approximate time requested to make their presentation on or before July 26, 2016. Time allotted for each presentation may be limited. If the number of registrants requesting to speak is greater than can be reasonably accommodated during the scheduled open public hearing session, FDA may conduct a lottery to determine the speakers for the scheduled open public hearing session. The contact person will notify interested persons regarding their request to speak by July 27, 2016.

Persons attending FDA's advisory committee meetings are advised that the Agency is not responsible for providing access to electrical outlets.

FDA welcomes the attendance of the public at its advisory committee meetings and will make every effort to accommodate persons with disabilities. If you require accommodations due to a disability, please contact AnnMarie Williams at AnnMarie.Williams@fda.hhs.gov or 301-796-5966 at least 7 days in advance of the meeting.

FDA is committed to the orderly conduct of its advisory committee meetings. Please visit our Web site at <http://www.fda.gov/AdvisoryCommittees/AboutAdvisoryCommittees/ucm111462.htm> for procedures on public conduct during advisory committee meetings.

Notice of this meeting is given under the Federal Advisory Committee Act (5 U.S.C. app. 2).

Dated: May 24, 2016.

Jill Hartzler Warner,

Associate Commissioner for Special Medical Programs.

[FR Doc. 2016-12658 Filed: 5/27/2016 8:45 am; Publication Date: 5/31/2016]