



DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

21 CFR Part 892

[Docket No. FDA-2026-N-4270]

Medical Devices; Radiology Devices; Classification of the Radiation Therapy Marking Device

AGENCY: Food and Drug Administration, HHS.

ACTION: Final amendment; final order.

SUMMARY: The Food and Drug Administration (FDA) is classifying the radiation therapy marking device into class II (special controls). The special controls that apply to the device type are identified in this order and will be part of the codified language for classification of the radiation therapy marking device. We are taking this action because we have determined that classifying the device into class II will provide a reasonable assurance of safety and effectiveness of the device. We believe this action will also enhance patients' access to beneficial innovative devices, in part by reducing regulatory burdens.

DATES: This order is effective [INSERT DATE OF PUBLICATION IN THE *FEDERAL REGISTER*]. The classification was applicable on December 10, 2021.

FOR FURTHER INFORMATION CONTACT: Lynne Fairobent, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 3663, Silver Spring, MD 20993-0002, 301-796-4817, Lynne.Fairobent@fda.hhs.gov.

SUPPLEMENTARY INFORMATION:

I. Background

Upon request, FDA (the Agency or we) has classified the radiation therapy marking device into class II (special controls), which we have determined will provide a reasonable assurance of safety and effectiveness of the device. In addition, we believe this action will

enhance patients' access to beneficial innovation, in part by reducing regulatory burdens by placing the device into a lower device class than the automatic class III assignment.

The automatic assignment of class III occurs by operation of law and without any action by FDA, regardless of the level of risk posed by the new device. Any device that was not in commercial distribution before May 28, 1976, is automatically classified into, and remains within, class III and requires premarket approval unless and until FDA takes an action to classify or reclassify the device (21 U.S.C. 360c(f)(1)). We refer to these devices as “postamendments devices” because they were not in commercial distribution prior to the date of enactment of the Medical Device Amendments of 1976, which amended the Federal Food, Drug, and Cosmetic Act (FD&C Act).

FDA may take a variety of actions in appropriate circumstances to classify or reclassify a device into class I or II. We may issue an order finding a new device to be substantially equivalent under section 513(i) of the FD&C Act (21 U.S.C. 360c(i)) to a predicate device that does not require premarket approval. We determine whether a new device is substantially equivalent to a predicate device by means of the procedures for premarket notification under section 510(k) of the FD&C Act (21 U.S.C. 360(k)) and part 807 (21 CFR part 807).

FDA may also classify a device through “De Novo” classification, a common name for the process authorized under section 513(f)(2) of the FD&C Act (see also part 860, subpart D (21 CFR part 860, subpart D)). Section 207 of the Food and Drug Administration Modernization Act of 1997 (Pub. L. 105-115) established the first procedure for De Novo classification. Section 607 of the Food and Drug Administration Safety and Innovation Act (Pub. L. 112-144) modified the De Novo classification process by adding a second procedure. A device sponsor may utilize either procedure for De Novo classification.

Under the first procedure, the person submits a premarket notification (510(k)) for a device that has not previously been classified. After receiving an order from FDA classifying the

device into class III under section 513(f)(1) of the FD&C Act, the person then requests a classification under section 513(f)(2).

Under the second procedure, rather than first submitting a 510(k) and then a request for classification, if the person determines that there is no legally marketed device upon which to base a determination of substantial equivalence, that person requests a classification under section 513(f)(2) of the FD&C Act.

Under either procedure for De Novo classification, FDA is required to classify the device by written order within 120 days. The classification will be according to the criteria under section 513(a)(1) of the FD&C Act. Although the device was automatically placed within class III, the De Novo classification is considered to be the initial classification of the device.

We believe this De Novo classification will enhance patients' access to beneficial innovation, in part by reducing regulatory burdens. When FDA classifies a device into class I or II via the De Novo process, the device can serve as a predicate for future devices of that type, including for 510(k)s (see section 513(f)(2)(B)(i) of the FD&C Act). As a result, other device sponsors do not have to submit a De Novo request or premarket approval application to market a substantially equivalent device (see section 513(i) of the FD&C Act, defining "substantial equivalence"). Instead, sponsors can use the less burdensome 510(k) process, when necessary, to market their device.

II. De Novo Classification

On June 22, 2020, FDA received Medical Precision B.V.'s request for De Novo classification of the Comfort Marker 2.0. FDA reviewed the request in order to classify the device under the criteria for classification set forth in section 513(a)(1) of the FD&C Act.

We classify devices into class II if general controls by themselves are insufficient to provide reasonable assurance of safety and effectiveness of the device, but there is sufficient information to establish special controls that, in combination with the general controls, provide reasonable assurance of the safety and effectiveness of the device for its intended use (see

section 513(a)(1)(B) of the FD&C Act). After review of the information submitted in the request, we determined that the device can be classified into class II with the establishment of special controls. FDA has determined that these special controls, in addition to the general controls, will provide reasonable assurance of the safety and effectiveness of the device.

Therefore, on December 10, 2021, FDA issued an order to the requester classifying the device into class II. In this final order, FDA is codifying the classification of the device by adding 21 CFR 892.5785.¹ We have named the generic type of device “radiation therapy marking device,” and it is identified as a powered device that transdermally delivers a permanent or temporary colorant to the skin for the purpose of placing marks to guide radiation therapy. This classification does not include devices with reusable or reprocessed needles or devices intended for diagnostic, therapeutic, or aesthetic use or to deliver other products for these uses.

FDA has identified the risks to health associated with this type of device and the measures required to mitigate these risks in table 1.

Table 1.--Risks to Health and Mitigation Measures for the Radiation Therapy Marking Device

Identified Risks to Health	Mitigation Measures
Adverse tissue reaction	Biocompatibility evaluation
Cross contamination and infection	Reprocessing validation; Sterilization validation; Non-clinical performance testing; Shelf-life testing; and Labeling
Needle stick injury to provider	Non-clinical performance testing; and Labeling
Device and/or software failure leading to ineffective marking	Clinical performance testing; Non-clinical performance testing; Software validation, verification, and hazard analysis; and Labeling
Electrical shock or electromagnetic interference with other devices	Electromagnetic compatibility testing; Electrical safety testing; and Labeling

¹ FDA notes that the “ACTION” caption for this final order is styled as “Final amendment; final order,” rather than “Final order.” Beginning in December 2019, this editorial change was made to indicate that the document “amends” the Code of Federal Regulations. The change was made in accordance with the Office of Federal Register’s (OFR) interpretations of the Federal Register Act (44 U.S.C. chapter 15), its implementing regulations (1 CFR 5.9 and parts 21 and 22), and the Document Drafting Handbook.

FDA has determined that special controls, in combination with the general controls, address these risks to health and provide reasonable assurance of safety and effectiveness of the device. For a device to fall within this classification, and thus avoid automatic classification in class III, it would have to comply with the special controls named in this final order. The necessary special controls appear in the regulation codified by this final order.

Under the FD&C Act, submission of a premarket notification under section 510(k) is required to reasonably assure the safety and effectiveness of class II devices unless FDA determines that the device type should be exempt under section 510(m) of the FD&C Act. At this time FDA has not made this determination for radiation therapy marking devices. This device is therefore subject to premarket notification requirements under section 510(k) of the FD&C Act.

III. Analysis of Environmental Impact

The Agency has determined under 21 CFR 25.34(b) that this action is of a type that does not normally have a significant effect on the human environment. Therefore, neither an environmental assessment nor an environmental impact statement is required.

IV. Paperwork Reduction Act of 1995

This final order establishes special controls that refer to previously approved collections of information found in other FDA regulations and guidance. These collections of information are subject to review by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of 1995 (44 U.S.C. 3501-3521). The collections of information in part 860, subpart D, regarding De Novo classification have been approved under OMB control number 0910-0844; the collections of information in 21 CFR part 814, subparts A through E, regarding premarket approval have been approved under OMB control number 0910-0231; the collections of information in part 807, subpart E, regarding premarket notification submissions have been approved under OMB control number 0910-0120; the collections of information in 21 CFR part 820 regarding quality management system regulation have been approved under OMB control

number 0910-0073; and the collections of information in 21 CFR part 801 regarding labeling have been approved under OMB control number 0910-0485.

List of Subjects in 21 CFR Part 892

Medical devices, Radiation protection, X-rays.

Therefore, under the Federal Food, Drug, and Cosmetic Act and under authority delegated to the Commissioner of Food and Drugs, 21 CFR part 892 is amended as follows:

PART 892--RADIOLOGY DEVICES

1. The authority citation for part 892 continues to read as follows:

Authority: 21 U.S.C. 351, 360, 360c, 360e, 360j, 360l, 371.

2. Add § 892.5785 to subpart F to read as follows:

§ 892.5785 Radiation therapy marking device.

(a) *Identification.* A radiation therapy marking device is a powered device that transdermally delivers a permanent or temporary colorant to the skin for the purpose of placing marks to guide radiation therapy. This classification does not include devices with reusable or reprocessed needles or devices intended for diagnostic, therapeutic, or aesthetic use or to deliver other products for these uses.

(b) *Classification.* Class II (special controls). The special controls for this device are:

(1) Design verification and validation must include:

(i) Documentation of performance data from studies that demonstrate:

(A) The indicated colorant is compatible with the device and its method of delivery;

(B) The device can reproducibly deliver the indicated colorant with the specifications described; and

(C) The length of time that compatible colorants remain visible on the skin following device application.

(ii) Documentation of performance data from studies that demonstrate:

(A) Accuracy and reproducibility of needle penetration depth;

- (B) Device protection from cross-contamination, including fluid ingress protection;
- (C) Adequacy of the cleaning and disinfection instructions to ensure that the reusable components of the device can be cleaned and disinfected; and
- (D) The sterility of all patient-contacting components (e.g., safety needle).
 - (iii) Documentation of performance data from studies that demonstrate electrical safety and electromagnetic compatibility of all electrical components of the device.
 - (iv) Documentation of performance data from studies that demonstrate continued sterility, package integrity, and device functionality over the intended shelf life.
 - (v) Documentation of software verification, validation, and hazard analysis.
- (2) The labeling required under § 801.109(c) of this chapter must include:
 - (i) An explanation of the device and the mechanism of operation;
 - (ii) Validated methods and instructions for reprocessing of any reusable components;
 - (iii) Disposal instructions; and
 - (iv) A shelf life for all sterile components.

Grace R. Graham,

Deputy Commissioner for Policy, Legislation, and International Affairs.

[FR Doc. 2026-08424 Filed: 4/29/2026 8:45 am; Publication Date: 4/30/2026]