



DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration

21 CFR Part 882

[Docket No. FDA-2026-N-6713]

Medical Devices; Neurological Devices; Classification of the External Lower Extremity Nerve Stimulator for Restless Legs Syndrome

AGENCY: Food and Drug Administration, HHS.

ACTION: Final amendment; final order.

SUMMARY: The Food and Drug Administration (FDA) is classifying the external lower extremity nerve stimulator for Restless Legs Syndrome into class II (special controls). The special controls that apply to the device type are identified in this order and will be part of the codified language for classification of the external lower extremity nerve stimulator for Restless Legs Syndrome. We are taking this action because we have determined that classifying the device into class II will provide a reasonable assurance of safety and effectiveness of the device. We believe this action will also enhance patients' access to beneficial innovative devices, in part by reducing regulatory burdens.

DATES: This order is effective [INSERT DATE OF PUBLICATION IN THE *FEDERAL REGISTER*]. The classification was applicable on April 17, 2023.

FOR FURTHER INFORMATION CONTACT: Chun Xu, Center for Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave., Bldg. 66, Rm. 4256, Silver Spring, MD 20993-0002, 301-796-2068, Chun.Xu@fda.hhs.gov.

SUPPLEMENTARY INFORMATION:

I. Background

Upon request, FDA (the Agency or we) has classified the external lower extremity nerve stimulator for Restless Legs Syndrome into class II (special controls), which we have determined

will provide a reasonable assurance of safety and effectiveness of the device. In addition, we believe this action will enhance patients' access to beneficial innovation, in part by reducing regulatory burdens by placing the device into a lower device class than the automatic class III assignment.

The automatic assignment of class III occurs by operation of law and without any action by FDA, regardless of the level of risk posed by the new device. Any device that was not in commercial distribution before May 28, 1976, is automatically classified into, and remains within, class III and requires premarket approval unless and until FDA takes an action to classify or reclassify the device (21 U.S.C. 360c(f)(1)). We refer to these devices as “postamendments devices” because they were not in commercial distribution prior to the date of enactment of the Medical Device Amendments of 1976, which amended the Federal Food, Drug, and Cosmetic Act (FD&C Act).

FDA may take a variety of actions in appropriate circumstances to classify or reclassify a device into class I or II. We may issue an order finding a new device to be substantially equivalent under section 513(i) of the FD&C Act (21 U.S.C. 360c(i)) to a predicate device that does not require premarket approval. We determine whether a new device is substantially equivalent to a predicate device by means of the procedures for premarket notification under section 510(k) of the FD&C Act (21 U.S.C. 360(k)) and part 807 (21 CFR part 807).

FDA may also classify a device through “De Novo” classification, a common name for the process authorized under section 513(f)(2) of the FD&C Act (see also part 860, subpart D (21 CFR part 860, subpart D)). Section 207 of the Food and Drug Administration Modernization Act of 1997 (Pub. L. 105-115) established the first procedure for De Novo classification. Section 607 of the Food and Drug Administration Safety and Innovation Act (Pub. L. 112-144) modified the De Novo classification process by adding a second procedure. A device sponsor may utilize either procedure for De Novo classification.

Under the first procedure, the person submits a premarket notification (510(k)) for a device that has not previously been classified. After receiving an order from FDA classifying the device into class III under section 513(f)(1) of the FD&C Act, the person then requests a classification under section 513(f)(2).

Under the second procedure, rather than first submitting a 510(k) and then a request for classification, if the person determines that there is no legally marketed device upon which to base a determination of substantial equivalence, that person requests a classification under section 513(f)(2) of the FD&C Act.

Under either procedure for De Novo classification, FDA is required to classify the device by written order within 120 days. The classification will be according to the criteria under section 513(a)(1) of the FD&C Act. Although the device was automatically placed within class III, the De Novo classification is considered to be the initial classification of the device.

We believe this De Novo classification will enhance patients' access to beneficial innovation, in part by reducing regulatory burdens. When FDA classifies a device into class I or II via the De Novo process, the device can serve as a predicate for future devices of that type, including for 510(k)s (see section 513(f)(2)(B)(i) of the FD&C Act). As a result, other device sponsors do not have to submit a De Novo request or premarket approval application to market a substantially equivalent device (see section 513(i) of the FD&C Act, defining "substantial equivalence"). Instead, sponsors can use the less burdensome 510(k) process, when necessary, to market their device.

II. De Novo Classification

On September 21, 2022, FDA received Noctrix Health, Inc.'s request for De Novo classification of the NTX100 Tonic Motor Activation (NTX100 ToMAc) System. FDA reviewed the request in order to classify the device under the criteria for classification set forth in section 513(a)(1) of the FD&C Act.

We classify devices into class II if general controls by themselves are insufficient to provide reasonable assurance of safety and effectiveness of the device, but there is sufficient information to establish special controls that, in combination with the general controls, provide reasonable assurance of the safety and effectiveness of the device for its intended use (see section 513(a)(1)(B) of the FD&C Act). After review of the information submitted in the request, we determined that the device can be classified into class II with the establishment of special controls. FDA has determined that these special controls, in addition to the general controls, will provide reasonable assurance of the safety and effectiveness of the device.

Therefore, on April 17, 2023, FDA issued an order to the requester classifying the device into class II. In this final order, FDA is codifying the classification of the device by adding 21 CFR 882.5887.¹ We have named the generic type of device “external lower extremity nerve stimulator for Restless Legs Syndrome,” and it is identified as a prescription device that uses external electrical stimulators and cutaneous electrodes to stimulate nerves in the lower extremity (e.g., peroneal nerves) and evoke tonic, sustained muscle activation in the legs to reduce the symptoms of Restless Legs Syndrome.

FDA has identified the risks to health associated with this type of device and the measures required to mitigate these risks in table 1.

Table 1.--Risks to Health and Mitigation Measures for External Lower Extremity Nerve Stimulator for Restless Legs Syndrome

Identified Risks to Health	Mitigation Measures
Adverse tissue reaction	Biocompatibility evaluation
Skin discomfort, burns, electrical shock, or pain at stimulation site	Electromagnetic compatibility testing; Electrical, mechanical, and thermal safety testing; Non-clinical performance testing; Software verification, validation, and hazard analysis; and Labeling
Worsening of Restless Legs Syndrome symptoms or disrupted sleep	Non-clinical performance testing; Software verification, validation, and hazard analysis; and Labeling

¹ FDA notes that the “ACTION” caption for this final order is styled as “Final amendment; final order,” rather than “Final order.” Beginning in December 2019, this editorial change was made to indicate that the document “amends” the Code of Federal Regulations. The change was made in accordance with the Office of Federal Register’s (OFR) interpretations of the Federal Register Act (44 U.S.C. chapter 15), its implementing regulations (1 CFR 5.9 and parts 21 and 22), and the Document Drafting Handbook.

User error or device failure due to interference with other devices, leading to delayed or ineffective treatment	Electromagnetic compatibility testing; Software verification, validation, and hazard analysis; and Labeling
--	---

FDA has determined that special controls, in combination with the general controls, address these risks to health and provide reasonable assurance of safety and effectiveness of the device. For a device to fall within this classification, and thus avoid automatic classification in class III, it would have to comply with the special controls named in this final order. The necessary special controls appear in the regulation codified by this final order.

At the time of classification, external lower extremity nerve stimulators for Restless Legs Syndrome are for prescription use only. Prescription devices are exempt from the requirement for adequate directions for use for the layperson under section 502(f)(1) of the FD&C Act (21 U.S.C. 352(f)(1)) and 21 CFR 801.5, as long as the conditions of 21 CFR 801.109 are met.

Under the FD&C Act, submission of a premarket notification under section 510(k) is required to reasonably assure the safety and effectiveness of class II devices unless FDA determines that the device type should be exempt under section 510(m) of the FD&C Act. At this time FDA has not made this determination for external lower extremity nerve stimulators for Restless Legs Syndrome. This device is therefore subject to premarket notification requirements under section 510(k) of the FD&C Act.

III. Analysis of Environmental Impact

The Agency has determined under 21 CFR 25.34(b) that this action is of a type that does not normally have a significant effect on the human environment. Therefore, neither an environmental assessment nor an environmental impact statement is required.

IV. Paperwork Reduction Act of 1995

This final order establishes special controls that refer to previously approved collections of information found in other FDA regulations and guidance. These collections of information are subject to review by the Office of Management and Budget (OMB) under the Paperwork

Reduction Act of 1995 (44 U.S.C. 3501-3521). The collections of information in part 860, subpart D, regarding De Novo classification have been approved under OMB control number 0910-0844; the collections of information in 21 CFR part 814, subparts A through E, regarding premarket approval have been approved under OMB control number 0910-0231; the collections of information in part 807, subpart E, regarding premarket notification submissions have been approved under OMB control number 0910-0120; the collections of information in 21 CFR part 820 regarding quality management system regulation have been approved under OMB control number 0910-0073; and the collections of information in 21 CFR part 801 regarding labeling have been approved under OMB control number 0910-0485.

List of Subjects in 21 CFR Part 882

Medical devices.

Therefore, under the Federal Food, Drug, and Cosmetic Act and under authority delegated to the Commissioner of Food and Drugs, 21 CFR part 882 is amended as follows:

PART 882—NEUROLOGICAL DEVICES

1. The authority citation for part 882 continues to read as follows:

Authority: 21 U.S.C. 351, 360, 360c, 360e, 360j, 360l, 371.

2. Add § 882.5887 to subpart F to read as follows:

§ 882.5887 External lower extremity nerve stimulator for Restless Legs Syndrome.

(a) *Identification.* An external lower extremity nerve stimulator for Restless Legs Syndrome is a prescription device that uses external electrical stimulators and cutaneous electrodes to stimulate nerves in the lower extremity (e.g., peroneal nerves) and evoke tonic, sustained muscle activation in the legs to reduce the symptoms of Restless Legs Syndrome.

(b) *Classification.* Class II (special controls). The special controls for this device are:

(1) Non-clinical performance testing must demonstrate that the device performs as intended under anticipated conditions of use. This testing must include:

(i) Characterization of the electrical stimulation parameters, including the following: waveforms; output modes; maximum output voltage and maximum output current (at 500 Ω , 2k Ω , and 10k Ω loads); pulse duration; frequency; net charge per pulse; and maximum phase charge, maximum current density, maximum average current, and maximum average power density (at 500 Ω);

(ii) Characterization of the therapy output across sudden and rapid changes in load impedance; and

(iii) Characterization of electrode performance, including electrical performance, adhesive integrity, shelf life, reusability, and variation of impedance over the use of therapy.

(2) The tissue-contacting components of the device must be demonstrated to be biocompatible.

(3) Performance testing must demonstrate electrical, thermal, and mechanical safety along with electromagnetic compatibility of the device in the intended use environment.

(4) Software verification, validation, and hazard analysis must be performed.

(5) Physician and patient labeling must include the following:

(i) Recommended treatment regimens, including frequency and duration of use, and identification of application site(s);

(ii) Typical sensations experienced during treatment;

(iii) Methods for identifying the appropriate stimulation intensity that is needed to reduce symptoms of Restless Legs Syndrome and is tolerable to patients;

(iv) A shelf life for the electrode and reuse information;

(v) Summaries of the electrical stimulation parameters and device technical parameters (including any wireless specifications); and

(vi) Instructions on how to maintain the device, including all user-interface components.

Grace R. Graham,

Deputy Commissioner for Policy, Legislation, and International Affairs.

